DE GRUYTER OPEN HUNGARIAN JOURNAL OF INDUSTRY AND CHEMISTRY Vol. 43(1) pp. 7–13 (2015) hjic.mk.uni-pannon.hu DOI: 10.1515/hjic-2015-0002



MULTI-ELEMENTAL INDUCTIVELY COUPLED PLASMA-OPTICAL EMISSION SPECTROSCOPIC CALIBRATION PROBLEMS OF THE SEQUENTIAL EXTRACTION PROCEDURE FOR THE FRACTIONATION OF THE HEAVY METAL CONTENT FROM AQUATIC SEDIMENTS

György Heltai, ¹* Ilona Fekete, ¹ Gábor Halász, ^{1,2} Katalin Kovács, ¹ Márk Horváth, ^{1,2} Anita Takács, ¹ Norbert Boros, ³ and Zoltán Győri²

- ¹ Department of Chemistry and Biochemistry, Szent István University, Pater K. u. 1., Gödöllő, 2100, HUNGARY
- ² Regional Knowledge Centre, Szent István University, Pater K. u. 1., Gödöllő, 2100, HUNGARY
- ³ Department of Environmental and Chemical Engineering, University of Debrecen, Ótemető u. 2-4., Debrecen, 4028, HUNGARY

For the characterisation of the environmental mobility of heavy metal contamination in aquatic sediments, the EU Bureau of Reference has proposed a fractionation by sequential extraction procedure. For its validation, the CRM-701 sample is available containing Cd, Cr, Cu, Ni, Pb, and Zn. In this paper, the matrix-matched calibration problems are presented. A multi-elemental inductively coupled plasma-optical emission technique is employed for the detection of heavy metals in the extracts. It was established that the sensitivities are strongly influenced by the extractants, which causes significant matrix effects: the sensitivities are strongly influenced by the solvents applied in extraction steps; the summarised recoveries show an acceptable agreement with the certified values; however, in the individual extraction steps for certain elements significant differences may occur due to the neglected interferences. Therefore, further optimisation is required utilising the flexible line selection possibility offered by the HORIBA Jobin Yvon ACTIVA-M instrument.

Keywords: fractionation, sequential extraction, multi-elemental ICP-OES, matrix-matched calibration, heavy metal contamination, sediments

1. Introduction

Environmentally toxic elements, such as heavy metals, are the most significant inorganic pollutants. The role of different chemical elements in food production has been an important question of agrochemical research since the middle of the nineteenth century. Numerous extraction methods have been developed for determining biologically available fractions of nutritionally relevant elements in soils. Environmental researchers have confronted similar problems over the last few decades of the twentieth century in characterising the human- and eco-toxicological impacts of increasing heavy metal pollution. This challenge has created new requirements for analytical methodology and contributed to the development of speciation analytical chemistry. This scientific discipline provides the foundation for the evaluation of toxicological impacts of chemical elemental species detectable in environmental systems [1,2]. In such procedures, the original physico-chemical structural forms of chemical elements present in the system must be preserved during sampling and sample preparation. Then, appropriate separation methods must be applied before element specific analytical detection (Fig.1).

The European Union established the European Virtual Institute for Speciation Analysis in 2002 for promoting cooperation, collection, reviewing, and discussion of information and knowledge related to chemical speciation analysis (www.speciation.net). Attention was drawn to the mobility and biological impact of heavy metals in environmental systems influenced by their chemical speciation. Biological risk assessment of contaminated areas therefore requires the total speciation of all chemical elements present in the system. However, due to the complex structure of solid environmental samples (e.g. soils, sediments, gravitation dust, mining waste, etc.), these analytical tasks in practice cannot be fully completed. As a compromise, the fractionation of the heavy metal content according to environmental mobility and estimation of biological availability by sequential extraction procedures have become the main research approaches. Papers published in this area and their citations exponentially increased in the period of 1995



Figure 1. Scheme of the speciation analysis procedure.

^{*}Correspondence: heltai.gyorgy@mkk.szie.hu



Figure 2. Strategy of serial single-step extractions.



Figure 3. Strategy of sequential extractions (S(R) stands for SOLVENT(REAGENT)).



Scheme 1. Element fractions in aquatic sediment as classified by Kersten and Förstner [4], Caroli [3], and Tessier [5] (left-hand side) and the original BCR sequential extraction steps [6-8] (right-hand side)

and 2013: from 200/year to 700/year; according to related citations from 1000/year to 19,000/year based on Thomson Reuters, Web of Science database. A primary risk assessment of heavy metal contamination in the soil/water/atmosphere system was based on the fractionation of metallic element content of environmental samples according to their environmental mobility. The term "fractionation" was recommended by the IUPAC in 2000 [2]. The mobility of an element according to the different solubilities of different binding classes can be estimated by solvents sequentially applied containing physico-chemically increasingly aggressive reagents [3,4]. In agricultural chemistry, the serial single step strategy (Fig.2), in environmental chemistry the strategy of sequential extraction has become dominant (Fig.3).

The first sequential extraction procedure for the characterisation of the mobility of heavy metals in aquatic sediments was developed by Tessier *et al.* [5], who classified the metal content of sediments according to the solubility and binding forms of metallic species as presented in *Scheme 1*.

Based on this essential work [5], numerous 5-8step sequential extraction schemes have been developed by which the original chemical structures (binding forms) are gradually dissolved and/or decomposed [9]. The complete duration of these procedures is 5–6 days, while the duration of the batch leaching steps is determined by a long setup time of solubility and heterogeneous chemical equilibria. The proposal for a 3(+1)-step sequential extraction procedure issued by BCR in 1993 is a simplified version of the above-mentioned schemes [6], but it is still time-consuming and labourintensive. As shown in *Scheme 1*, the water soluble, weakly adsorbed, and carbonate-associated metal fractions are extracted simultaneously by acetic acid in the first leaching step. In the second step, the metal fractions associated with reducible Fe and Mn oxides are produced after reduction by hydroxylamine. A fraction of metals coordinated to organic molecules and associated with sulphides is extracted after oxidation by H_2O_2 in the third step. In the extra step, the residual fraction can be obtained by means of digestion with *aqua regia* or HNO₃/H₂O₂.

Re-adsorption of extracted metals during phase separation can lead to analyte losses. In the extracts, the high concentration of the reagent may cause strong during spectrochemical matrix effects element detection. The CRM 601 sediment reference material issued in 1997 has been certified for this procedure only for Cd, Cr, Ni, Pb, and Zn [7,8]. During the analysis of the BCR 701 sediment materials, the original BCR procedure was modified to avoid analyte loss during the second leaching step due to pH instability [10–13]. However, this modification in BCR procedure may cause significant differences relative to previously obtained data. The BCR procedures in their original and modified versions were successfully applied to aquatic sediments and soils; however, soil CRM has not been issued for this procedure up to now [14-21].

It can be concluded that the Tessier-based fractionation procedures [5] are widely used for heavy metal contamination assessments in soils and aquatic sediments, but the realistic interpretation of these data requires standardisation. However, this has been hindered due to methodological problems of measurements and analysis [22–23]. These problems limiting the application of the BCR sequential

Table 1. Overview of the cooperative research projects between Szent István University Gödöllő and the Technical University Košice

Limitations of the	Achievements by SZIE and TU Košice
BCR scheme	research groups
Does not match the – natural mobilisation pathways	→ Better modelling of environmental mobility by the application of H ₂ O and H ₂ O/CO ₂ solvents [24].
Does not preserve – the intact species of the sample	→ Application of aqueous extractions for the elaboration of biological tests [25].
Time consuming – (4-5 days) –	 Acceleration of batch leaching by means of sonification [26]. Development of continuous flow operation in the supercritical fluid extractor for CO₂, H₂O and H₂O/CO₂ solvents [24]
Cross-contamination /	
during operation and -	 Reduction of operations and reagents
applied reagent	
Limited applicability and validation:	
Validation and – available CRM only for sediments	 Extension of the application to other matrices: soils, gravitation dusts, biofilms, red mud, and composts [26-30].
Validation of the – procedure for Cd,	→ Multi-elemental calibration and validation for different matrices and extracts [27].
Cr, Cu, Ni, Pb, Zn High reagent and <i>r</i> organic content of the extracts	
LIE EALIAULS	

extraction procedure have initiated a cooperative research project between Szent István University Gödöllő and the Technical University Košice as summarised in *Table 1*.

The aim of the given study was the evaluation of changes in the environmental mobility of heavy metals at two polluted areas (Upper Tisza/2000-2001 and Gödöllő-Isaszeg lake system/1995-1998) by means of the repeated BCR fractionation of heavy metals in the sediments and soils [31-34]. A flexible, multi-elemental Inductively Coupled Plasma-Optical Emission (ICP-OES) technique with a Charge-Coupled Device (CCD)detection system was employed instead of the formerly applied flame atomic absorption spectroscopy (FAAS). The latter had to be optimised for each analyte element individually. Hereby, the problems of the matrix-matched multi-elemental calibration of the ICP-OES method are described using different solvent applied from the BCR procedure. The validation of the calibration was controlled by the fractionation of the CRM 701 sample.

2. Experimental

2.1. Elemental Analysis and Instrumentation

For element detection in different extractant matrices, the HORIBA Jobin Yvon ACTIVA-M ICP-OE spectrometer was employed, which is able to record the entire UV-vis spectrum range by rapid scanning CCDdetection. It also makes possible the flexible analytical line selection for 75 elements. The wavelength scale calibration of recorded spectra is based on a reference line position set by the user. The C 193.026 nm line is proposed by the manufacturer in routine analytical applications. The instrument's scheme and technical



Figure 4. Scheme of the HORIBA Jobin Yvon ACTIVA-M ICP-OE spectrometer and relevant technical parameters

parameters are presented in Fig.4. Introduction of the sample can be performed by a concentric glass nebuliser and glass cyclone spray chamber fed by a 3-channel peristaltic pump.

2.2. ICP-OES Operation Parameters

The operation parameters of ICP-OES were selected on the basis of the manufacturer's manual proposed for multi-elemental analysis as follows:

- Incident power: 1000 W,
- Plasma gas flow: 13.14 dm³ min⁻¹,
- Sheath gas flow: 0.32 dm³ min⁻¹,
- Nebuliser gas flow: 0.28 dm³ min⁻¹,
- Nebuliser gas pressure: 2.86 bar, and
- Nebuliser sample flow uptake: 0.85 cm³ min⁻¹.

2.3. Matrix Matched Calibration Solutions

Amounts of 1 mg dm⁻³ (in cc HNO₃) mono-elemental Cd, Co, Cr, Cu, Fe, Mn, Ni, Pb, V, and Zn (Précis Certified Standards for ICP Spectroscopy, HORIBA Scientific, France) standard solutions corresponding to 0, 5, and 20 mg dm⁻³ concentrations were pipetted into a 50 cm³ volumetric flask and then diluted by solvents applied in the BCR sequential extraction procedure:

- 1st step: 0.11 M acetic acid.
- 2nd step: 0.1 M hydroxylamine-hydrochloride.
- 3rd step: 1 M ammonium acetate.
- extra step: 13 g/g nitric acid.

2.4. Analytical Lines

The following analytical lines were selected according to the manufacturer's proposal for multi-elemental analysis:

- Cd 214.441 nm
- Co 230.785 nm
- Cr 205.571 nm
- Cu 324.754 nm
- Fe 259.940 nm and 240.489 nm
- Mn 257.610 nm
- Ni 222.296 nm (instead of 221.650 nm!)
- Pb 220.353 nm
- V 309.311 nm
- Zn 213.857 nm

In the first, second, and extra step, the wavelength readout started from the C 193.026 nm reference line. In the third step, the correct wavelength readout was only

Table 2. The calibration lines (y = mx+b, at wave length λ in nm), correlation R^2 values, residual standard deviations (RSD), background equivalent concentrations (BEC in µg dm⁻³), and detection limits ($C_{L(3\sigma)}$ in µg dm⁻³) obtained by a 0.11 M acetic acid matrix.

	λ	$C_{L(3\sigma)}$	BEC	m	b	R^2	RSD
Cd	214.441	0.5	65	43778	2867	0.9999	1.1%
Co	230.785	0.8	145	41860	6064.7	0.9997	2.5%
Cr	205.571	1.5	174	18298	3184.2	0.9996	2.9%
Cu	324.754	0.2	121	159951	19374	0.9998	2.0%
Fe	259.940	0.5	72	205720	14716	0.9999	1.2%
Mn	257.610	0.2	105	902188	94673	0.9999	1.8%
Ni	222.296	2.0	115	7154	823.38	0.9998	2.0%
Pb	220.353	3.5	17	4087.4	68.946	1.0000	0.3%
V	309.311	0.6	34	259817	8897	1.0000	0.5%
Zn	213.857	0.6	148	60540	8951	0.9998	2.2%

Table 4. The calibration lines (y = mx+b, at wave length λ in nm), correlation R^2 values, residual standard deviations (RSD), background equivalent concentrations (BEC in µg dm⁻³), and detection limits ($C_{L(3\sigma)}$ in µg dm⁻³) obtained by a 1.00 M ammoniumacetate matrix.

	λ	$C_{L(3\sigma)}$	BEC	m	b	R^2	RSD
Cd	214.441	0.6	57	70051	-4008.3	1.0000	1.0%
Co	230.785	1.4	2	54840	-119.26	1.0000	0.03%
Cr	205.571	1.2	21	25579	546.52	1.0000	0.3%
Cu	324.754	2.5	213	151053	32156	0.9994	3.5%
Fe	259.940	15.8	160	258243	41379	1.0000	0.6%
Mn	257.610	0.4	2	1291505	2013	1.0000	0.02%
Ni	222.296	2.3	17	9902	-170.25	1.0000	0.3%
Pb	220.353	3.5	148	5905.2	-874.08	0.9997	2.6%
V	309.311	1.4	85	312207	26636	0.9999	1.4%
Zn	213.857	14.5	33	79775	2643.2	0.9968	1.9%

Table 5. The calibration lines (y = mx+b, at wave length λ in nm), correlation R^2 values, residual standard deviations (RSD), background equivalent concentrations (BEC in µg dm⁻³), and detection limits ($C_{L(30)}$ in µg dm⁻³) obtained by a 13 g/g nitric acid and hydrogen peroxide matrix.

	λ	$C_{L(3\sigma)}$	BEC	т	b	R^2	RSD
Cd	214.441	0.4	392	74770	29293	0.9980	6.6%
Co	230.785	3.5	319	50151	16009	0.9986	5.4%
Cr	205.571	1.7	328	23088	7580.7	0.9986	5.5%
Cu	324.754	0.9	343	104399	35859	0.9985	5.6%
Fe	259.940	0.7	219	216961	47594	0.9995	3.2%
Mn	257.610	0.5	211	1056725	222941	1.0000	3.6%
Ni	222.296	0.5	320	9282.3	2974.7	0.9987	5.4%
Pb	220.353	8.7	265	6483.7	1717.4	0.9991	4.4%
V	309.311	0.7	54	256488	13775	1.0000	0.9%
Zn	213.857	17.3	517	87624	45308	0.9968	8.3%

possible when the readout started from the Ar 404.442 nm reference line due to the use of 1 M ammonium acetate solution.

2.5. Validation

For validation of the CRM 701 sediment, the reference sample was applied. The fractionation of its element content was performed according to the modified BCR proposal [10–13] in triplicates. For the extraction of the

Hungarian Journal of Industry and Chemistry

Table 3. The calibration lines (y = mx+b, at wave length λ in nm), correlation R^2 values, residual standard deviations (RSD), background equivalent concentrations (BEC in µg dm⁻³), and detection limits ($C_{L(3\sigma)}$ in µg dm⁻³) obtained by a 0.10 M hydroxylamine-hydrochloride matrix.

	λ	$C_{L(3\sigma)}$	BEC	т	b	R^2	RSD
Cd	214.441	0.9	87	47809	-4145.2	0.9999	1.5%
Co	230.785	1.4	19	44538	-856.49	1.0000	0.3%
Cr	205.571	1.4	14	18911	-258.4	1.0000	0.2%
Cu	324.754	0.9	72	163871	11765	1.0000	1.0%
Fe	259.940	0.3	1	204879	236.82	1.0000	0.2%
Mn	257.610	0.2	63	911472	57160	0.9999	1.1%
Ni	222.296	3.7	63	7505.8	471.04	1.0000	1.0%
Pb	220.353	3.9	4	4375.2	-17.636	1.0000	0.2%
V	309.311	0.2	58	260057	-15164	1.0000	1.0%
Zn	213.857	1.5	45	63846	2845.5	1.0000	0.5%

Table 6. Results for the BCR-fractionation of element content (in μ g dm⁻³) of the CRM 701 sample.

steps		measured	certified	recovery
	Cd	6.45 ± 0.07	7.34±0.35	88%
	Co	1.68 ± 0.06		
	Cr	1.88 ± 0.08	2.26 ± 0.16	83%
	Cu	44.93 ± 0.25	49.3±1.7	91%
1 st	Fe	57.74 ± 2.53		
1	Mn	133.17 ± 6.52		
	Ni	12.52 ± 0.34	15.4±0.9	81%
	Pb	1.92 ± 0.19	3.18 ± 0.21	60%
	V	0.06 ± 0.23		
	Zn	160.07 ± 4.22	205±6	78%
	Cd	3.96 ± 0.08	3.77±0.28	105%
	Co	2.71 ± 0.03		
	Cr	26.53 ± 0.78	45.7±2	58%
	Cu	87.61 ± 0.81	124±3	71%
and	Fe	5648 ± 50		
2.4	Mn	106.99 ± 2.90		
	Ni	19.55 ± 0.35	26.6±1.3	73%
	Pb	92.17 ± 2.24	126±3	73%
	V	8.50 ± 0.09		
	Zn	98.77 ± 2.27	114±5	87%
	Cd	1.17 ± 0.01	0.27±0.06	435%
	Co	1.80 ± 0.08		
	Cr	196.4± 8.24	143±7	137%
	Cu	63.59 ± 2.92	55.2±4	115%
ard	Fe	4652 ± 133		
3	Mn	43.65 ± 1.25		
	Ni	24.49 ± 0.35	15.3±0.9	160%
	Pb	60.14 ± 0.82	9.3±2	647%
	V	5.86 ± 0.21		
	Zn	66.44 ± 0.62	45.7±4	145%
	Cd	1.38 ± 0.01		
	Co	7.63 ± 0.18		
	Cr	101.12 ± 4.79		
extra	Cu	49.48 ± 1.92		
	Fe	24988 ± 805		
	Mn	302.9 ± 3.7		
	Ni	44.50 ± 1.36		
	Pb	12.77 ± 0.30		
	V	40.30 ± 0.67		
	Zn	127.12 ± 1.36		

residual fraction in the extra step and for the determination of the pseudo total element content microwave supported cc HNO_3/H_2O_2 digestion was applied.

Table 7. Comparison of the summarised element content (in mg kg⁻¹) extracted in BCR steps 1-3 using certified values for the CRM 701 sample.

	steps 1-3	certified	recovery
Cd	11.58	11.38	102%
Co	6.19		
Cr	224.8	191.0	118%
Cu	196.1	228.5	86%
Fe	10358		
Mn	283.8		
Ni	56.6	57.3	99%
Pb	154.2	138.5	111%
V	14.42		
Zn	325.3	364.7	89%

3. Results and Analysis

The parameters of analytical calibration lines, background equivalent concentrations (BEC) and detection limits ($C_{L(3\sigma)}$) obtained by different extraction matrices are summarised in *Tables 2–5*. It can be established that the extraction matrix solvents significantly influence the sensitivity (*B*) values for all elements in the order of $B(3^{rd}) > B(extra) > B(2^{rd})$ > B(1st) as a function of steps taken. With the exception of some outliers, the residual standard deviation values are in the range of 1–2% and the detection limits are in the range of 0.5–2 µg dm⁻³. The outliers for Pb, Zn and Fe in certain matrices refer to neglected spectral interferences.

The results of the fractionation of the element content of the CRM 701 sample by applying the matrix matched calibration for determining the element contents in the extracts are summarised and compared with the certified and pseudo-total values in *Tables 6–8*. It can be seen that the metal content in BCR Steps 1–3 shows an acceptable agreement (< mg dm⁻³) with certified values. In the individual steps, there are larger differences for elements than are acceptable: 1^{st} step: Pb; 2^{nd} step: Cr, Cu, and Pb; 3^{rd} step: Cd, Ni, Pb, and Zn.

4. Conclusion

Based on the results presented here, it can be concluded that the extractants applied in the BCR sequential extraction procedure cause significant matrix effects even in the case of matrix-matched multi-elemental calibration of the ICP-OES determination of element content. The sensitivities were strongly influenced by the extractant media, therefore, a much higher incident power was required to eliminate these differences to ensure conditions that are more robust. The recoveries of certified values of the CRM 701 sample were within the acceptable range when the summary of extracted amounts in the BCR Steps 1-3 was compared. Significant differences occurring in individual extraction steps between the measured and certified values refer to neglected spectral interferences. This requires a detailed study utilising the flexible line

Table 8. Comparison of the summarised element content (in mg kg⁻¹) extracted in BCR steps 1-3 using pseudo-total values.

	steps 1-3+extra	pseudo-total	recovery
Cd	12.96	9.63	135%
Co	13.81	10.68	129%
Cr	325.9	290.5	112%
Cu	245.6	257.7	95%
Fe	35323	31285	113%
Mn	586.4	567.7	103%
Ni	101.0	77.2	131%
Pb	167.0	158.2	106%
V	54.68	39.20	140%
Zn	452.3	417.7	108%

selection possibility offered by the HORIBA Jobin Yvon ACTIVA-M instrument.

Acknowledgement

This work was supported by the Hungarian Scientific Research Fund (OTKA 108558), *KTIA* AIK-12-1-2012-0012 and the Research Centre of Excellence - 8526-5/2014/TUDPOL.

REFERENCES

- Gunn, A.M.; Winnard, D.A.; Hunt, D.T.E.: Trace metal speciation in sediment and soils, in *Metal speciation: Theory, analysis and application*, Eds.: Kramer, J.R.; Allen, H.E. (Lewis Publications, Boca Raton, FL, USA) 1988 ISBN 0873711408 9780873711401
- [2] Templeton, D.M.; Ariese, F.; Cornelis, R.; Danielsson, L.G.; Muntau, H.; van Leeuwen, H.P.; Lobinski, R.: Guidelines for terms related to chemical speciation and fractionation of elements. Definitions, structural aspects, and methodological approaches, *Pure Appl. Chem.*, 2000 72, 1453–1470 10.1351/pac200072081453
- [3] Caroli, S.: Element speciation in bioorganic chemistry (Wiley, New York, USA) 1996 ISBN 0-471-57641-7
- [4] Kersten, M.; Förstner, M.: Speciation of trace metals in sediments and combustion waste. *In Chemical speciation in the environment*, Eds.: Ure, A.M.; Davidson, C.M. (Blackie Academic and Professional, London, U.K.) 1995, pp. 237-275
- [5] Tessier, A.; Campbell, P.G.C.; Bisson, M.: Sequential extraction procedure for the speciation of particulate trace metals, *Anal. Chem.*, 1979, **51**, 844–851 10.1021/ac50043a017
- [6] Ure, A.M.; Quevauviller, Ph.; Muntau, H.; Griepink, B.: Speciation of heavy-metals in soils and sediments - an account of the improvement and harmonization of extraction techniques undertaken under the auspices of the BCR of the Commission of the European Communities, *Int. J. Envir. Anal. Chem.*, 1993 **51**, 135–151 10.1080/0306731908027619

- [7] Quevauviller, P.; Rauret, G.; López-Sánchez, J.F.; Rubio, R.; Ure, A.; Muntau, H.: Certification of trace metal extractable contents in a sediment reference material (CRM 601) following a threestep sequential extraction procedure, *Sci. Total Environ.*, 1997 **205**, 223–234 10.1016/S0048-9697(97)00205-2
- [8] López-Sánchez, J.F.; Sahuquillo, A.; Fiedler, H.D.; Rubio, R.; Rauret, G.; Muntau, H.; Quevauviller, Ph.: CRM 601, a stable material for its extractable content of heavy metals, *Analyst*, 1998 **123**, 1675– 1677 10.1039/A8022720J
- [9] Gleyzes, C.; Tellier, S.; Astruc, M.: Fractionation studies of trace elements in contaminated soils and sediments: a review of sequential extraction procedures, *Trends Anal. Chem.*, 2002 21, 451-467 10.1016/S0165-9936(02)00603-9
- [10] Rauret, G.; López-Sánchez, J.F.; Sahuquillo, A.; Rubio, R.; Davidson, C.; Ure, A.; Quevauviller, Ph.: Improvement of the BCR three step sequential extraction procedure prior to the certification of new sediment and soil reference materials, J. Envir. Monit., 1999 1, 57–61 10.1039/A807854H
- [11] Sahuquillo, A.; López-Sánchez, J.F.; Rubio, R.; Rauret, G.; Thomas, R.P.; Davidson, C.M.; Ure, A.M.: Use of a certified reference material for extractable trace metals to assess sources of uncertainty in the BCR three-stage sequential extraction procedure, *Anal. Chim. Acta*, 1999 **382**, 317–327 10.1016/S0003-2670(98)00754-5
- [12] Pueyo, M.; Rauret, G.; Lück, D.; Yli-Halla, M.; Muntau, H.; Quevauviller, Ph.; López-Sánchez, J.F.: Certification of the extractable contents of Cd, Cr, Cu, Ni, Pb and Zn in a freshwater sediment following a collaboratively tested and optimised three-step sequential extraction procedure, *J. Envir. Monit.*, 2001 **3**, 243–250 10.1039/b01025k
- [13] Rauret, G.; López-Sánchez, J.F.; Lück, D.; Yli-Halla, M.; Muntau, H.; Quevauviller, Ph.: Certification of the extractable contents of Cd, Cr, Cu, Ni, Pb and Zn in a freshwater sediment following a collaboratively tested and optimised three-step sequential extraction procedure, European Commission BCR information reference materials BCR 701, 2001 ISBN 92-894-0755-7
- [14] Sahuquillo, A.; Rigol, A.; Rauret, G.: Overview of the use of leaching/extraction tests for risk assessment of trace metals in contaminated soils and sediments, *Trends Anal. Chem.*, 2003 22, 152– 159 10.1016/S0165-9936(03)00303-0
- [15] Žemberyová, M.; Barteková, J.; Hagarová, I.: The utilization of modified BCR three-step sequential extraction procedure for the fractionation of Cd, Cr, Cu, Ni, Pb and Zn in soil reference materials of different origins, *Talanta*, 2006 **70**, 973–978 10.1016/j.talanta.2006.05.057
- [16] Kubová, J.; Streško, V.; Bujdoš, M.; Matúš, P.; Medved, J.: Fractionation of various elements in CRMs and in polluted soils, *Anal. Bioanal. Chem.*, 2004 **379**, 108–114 10.1007/s00214-004-2505-5

- [17] Kubová, J.; Matúš, P.; Hagarová, I.; Bujdoš, M.; Medved, J.: Utilization of optimised BCR threestep sequential and dilute HCl single extraction procedures for soil-plant metal transfer predictions in contaminated lands, *Talanta*, 2008 **75**, 1110– 1122 10.1016/j.talanta.2008.01.002
- [18] Konradi, E.A.; Frentiu, T.; Ponta, M.; Cordos, E.: Use of sequential extraction to assess metal fractionation in soils from Bozanta Mare, Romania, *Acta Univers. Cibiniensis Ser. F Chemia*, 2005 8(2), 5–12 http://saiapm.ulbsibiu.ro/rom/cercetare/art/art2.pdf
- [19] Okoro, H.K.; Fatoki, O.S.; Adekola, F.A.; Ximba, B.J.; Snyman, R.G.: A review of sequential extraction procedures for heavy metals speciation in soil and sediments, *Open Access Sci. Rep.*, 2012 1(3): 181 10.4172/scientificreports.181
- [20] Jena, V.; Gupta, S.; Dhundhel, R.S.; Matic, N.; Frančišković Bilinski, S.; Devic, N.: Determination of total heavy metal by sequential extraction from soil. *Int. J. Res. Environ. Sci. Technol.*, 2013 3(1), 35–38 ISSN 2249-9695
- [21] Zimmerman, A.J.; Weindorf, D.C.: Heavy metal and trace metal analysis in soil by sequential extraction: A review of procedures. Int. J. Anal. Chem., 2010 7, Article ID 387803 10.1155/2010/387803
- [22] Zimmerman, A.J.; Weindorf, D.C.: Heavy metal and trace metal analysis in soil by sequential extraction: A review of procedures, *Int. J. Anal. Chem.*, 2010 7, 387803 10.1155/2010/387803
- [23] Dabek-Zlotorzynska, E.; Kelly, M.; Chen, H.; Chakrabarti, C.L.: Application of capillary electrophoresis combined with a modified BCR sequential extraction for estimating of distribution of selected trace metals in $PM_{2.5}$ fractions of urban airborne particulate matter, *Chemosphere*, 2005 **58**, 1365–1376 10.1016/j.chemosphere.2004.09.082
- [24] Heltai, Gy.; Fehér, B.; Percsich, K.; Barabás, B.; Fekete, I.: Application of sequential extraction with supercritical CO₂, subcritical H₂O, and an H₂O/CO₂ mixture for estimation of environmentally mobile heavy metal fractions in sediments, *Anal. Bioanal. Chem.*, 2002 **373**(8), 863–866 10.1007/s00216-002-1388-6
- [25] Heltai, Gy.: Chemical interpretation of a new sequential extraction scheme based on supercritical CO₂ and subcritical H₂O solvents, *Bull. Szent István University*, 2005, 107–122
- [26] Rusnák, R.; Halász, G.; Horváth, M.; Remeteiová, D.: Preliminary results on the intensification of the BCR sequential extraction with sonication for sediments, soils, and gravitation dust sediment samples, *Toxicol. Environment. Chem.*, 2010 **92**(3), 443–452 10.1080/02772240903300139
- [27] Horváth, M.; Halász, G.; Kucanová, E.; Kuciková, B.; Fekete, I.; Remeteiová, D.; Heltai, Gy.; Flórián, K.: Sequential extraction studies on aquatic sediment and biofilm samples for the assessment of heavy metal mobility, *Microchem. J.*, 2013 **107**(3), 121–125 10.1016/j.microc.2012.05.024

- [28] Flórián, K.; Matherny, M.; Nickel, H.; Pliešovská, N.; Uhrinová, K.: Environmental characteristics of the atmosphere of residential agglomerations - I. Dustiness of the agglomerations, *Chem. Pap.*, 2003 57(5), 369–373
- [29] Flórián, K.; Matherny, M.; Nickel, H.; Pliešovská, N.; Uhrinová, K.: Environmental characteristics of the atmosphere of residential agglomerations - II. Main, minor, and trace elements in the gravitation dust sediments, *Chem. Pap.*, 2003 57(5), 374–381
- [30] Remeteiová, D.; Sminčáková, E.; Flórián, K.: Study of the chemical properties of gravitation dust sediments, *Microchim. Acta*, 2007 **156**(1-2), 109– 113 10.1007/s00604-006-0603-0
- [31] Shiowatana, J.; Tantidanai, N.; Noobkaew, S.; Nacapricha, D.: A novel continuous-flow sequential extraction procedure for metal speciation in solids, *J. Environ. Qual.*, 2001 **30**, 1195–1205 10.2134/jeq2001.3041195x

- [32] Heltai, Gy.; Fekete, I.; Gémesi, Z.; Percsich, K.; Flórián, K.; Tarr, Zs. : Environmental evaluation of a local lake chain affected by wastewater by means of spectrochemical analytical methods, *Microchem*. *J.*, 1998 **59**, 125–135 10.1006/mchj.1998.1574
- [33] Győri, Z.; Alapi, K.; Szilágyi, Sz.: Effects of heavy metal on floodplain soils and pastures of the River Tisza, Hungary I. In natural attenuation of metals along the Tisza river–floodplain–wetlands continuum (University of Debrecen, Debrecen, Hungary) 2003, pp. 146–160 ISBN: 963 472 726 3
- [34] Győri, Z.; Alapi, K.; Sipos, P.; Zubor, Á.: Effects of heavy metal on floodplain soils and pastures of the River Tisza, Hungary II. In natural attenuation of metals along the Tisza river-floodplainwetlands continuum (University of Debrecen, Debrecen, Hungary) 2003, pp. 161–163 ISBN: 963 472 726 3







About water for water – WATER AND WASTEWATER TREATMENT SYSTEM OPERATING ENGINEER postgraduate specialist training course with the co-operation of one of Hungary's oldest historical universities and a company of international reputation

"WATER IS ESSENTIAL FOR LIFE ON EARTH AS MUCH AS GOOD QUALITY EDUCATION IS FOR PROFESSIONAL EXPERTS."

The <u>University of Pannonia Faculty of Engineering</u> and the <u>Nagykanizsa Campus</u> along with <u>HIDROFILT Water and</u> <u>Wastewater Treatment Ltd.</u> are offering a postgraduate course programme in water and wastewater treatment.

Hungary has long been recognised as a great water power with a historical background in the field of water treatment and purification. The University of Pannonia has a long tradition of academic excellence based on the training of engineering specialists for industry and the economy. The Faculty of Engineering and the Nagykanizsa Campus of the University of Pannonia are committed to forging partnerships with industry. Our aim is to create a generation of engineers that will take a lead in the development of innovative and sustainable technologies.

THE WATER AND WASTEWATER TREATMENT SYSTEM OPERATION PROGRAMME is designed for those engineering and non-engineering graduates who wish to pursue a career in water and wastewater treatment.

The Nagykanizsa Campus offers a wide spectrum of quality programmes that not only enhance students' professional development, but also familiarise them with the natural and historic values of Hungary. Mentors will assist students in organising their educational programmes and review their theses. Each student benefits from the opportunity of acquiring a broad range of transferable skills towards their future career within an outstanding academic environment.

Duration of the course programme: 2 semesters/4 months.

Theoretical training in water and wastewater treatment technologies is provided by the University of Pannonia, and Hidrofilt Ltd. Practical work and laboratory sessions will be carried out with the help of professional mentors from Hidrofilt Ltd. The Faculty of Engineering provides course materials and notes that are included in the package. Competencies to be acquired:

- Classification and adequate employment of the theoretical knowledge and practical methods acquired in the field; independent decision-making and development skills.
- Ability to analyse the qualitative and quantitative characteristics of water and wastewater; creating, implementing and assessing measurement plans.
- Optimal employment and management of state-of-the-art water and wastewater treatment technologies.
- Operating, planning and managing water supply and water and wastewater treatment technologies.
- Knowledge and employment of quality assurance and technical control.
- Analytical skills in technical, economical, and social relations in water and wastewater treatment.
- Evaluation of conceptions and programmes regarding water on national and international levels.

The qualification provides skills and abilities that can be employed profitably not only on a national but also on an international level. Graduates are able to design and implement investments related to water and wastewater treatment and maintain systems in operation.

We offer a practice-oriented training programme in addition to high-quality academic education. Those participating in the study programme gain state-of-the-art knowledge that fits the needs of the market and acquire the most advanced water purification technologies applied in the industry.